



Product Service

# CERTIFICATE

No. Z1 16 09 22430 044

**Holder of Certificate: Alu Rehab AS**Bedriftsvegen 23  
4353 Klepp Stasjon  
NORWAY**Certification Mark:****Product:****Care wheelchairs**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713077935

Valid until: 2021-09-14

Date, 2016-09-15  (Torsten Zimmer)

Page 1 of 2



Product Service

## CERTIFICATE

No. Z1 16 09 22430 044

**Model(s):** **Netti Vision**

### Parameters:

total length incl. footrest:	990 mm
total length without footrest:	700 mm
width:	550 mm
height:	765 - 980 mm
seat width:	400 mm
seat depth:	350 - 425 mm
seat height:	380 mm
backrest height:	455 mm
Seat plane angle:	-2° - 30°
backrest angle:	90° - 110°
max. user weight:	125 kg
brake system:	Drum brake (assistant brake)
back wheels:	
-size	16 x 3/8"
-tyres	solid rubber (puncture proof)
-adjustable	no
-fixation	fix
front wheels:	
-size	ø150 x 27 mm
-tyres	solid rubber (puncture proof)
-adjustable	yes
-fixation	castor with quick release
frame material:	aluminum
armrest:	height adjustable
legrest:	height adjustable, angle adjustable and removable
anti tip device:	adjustable

**Tested according to:** EN 12183:2014

**Production Facility(ies):** 57370