



Product Service

CERTIFICATE

No. Z1 12 01 22430 032

Holder of Certificate: **Alu Rehab AS**

Bedriftsveien 23
4353 Klepp Stasjon
NORWAY

Production Facility(ies):
57370

Certification Mark:



Product: **Electrical wheelchairs class B**

Model(s): **Netti Mobile**

Parameters:

total height:	1010 mm
total length:	1110 mm
total width:	600 mm
obstacle height:	100 mm
max. speed:	10 km/h
max. load:	140 kg
max. safe slope:	6° (in the standard seat position as specified by the manufacturer)
driving motor:	24 VDC, 320 W
motor for the adjustment of the seat angle:	24 VDC, max. 2.5 A
motor for adjustable legrests:	24 VDC, max. 1.5 A
motor for adjustable backrest:	24 VDC, max. 2.5 A
power electronics:	Dynamic Controls, type: DX2-PMA90L + DX2-ACT4 or type: DK-PMB31 (SHARK)
batteries:	2x (12 VDC, 50 Ah (C20))
battery charger:	IN: 100-240 VAC; 50/60 Hz; 3.5 A; Class II OUT: 24 VDC, 5 A

Tested according to: EN 12184:2009

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 71397806 Revision 1

Date, 2012-02-20


(Torsten Zimmer)

Page 1 of 1



ZERTIFIKAT ♦ CERTIFICATE ♦ CERTIFICADO ♦ CERTIFICAT ♦ СЕРТИФИКАТ ♦ 認証証書 ♦ CERTIFICATE ♦ ZERTIFIKAT